



Nederlands Meetinstituut

EC type-approval certificate

Number **T2170** Revision 1
Project number 10024223 Page
1 of 4

Issued by NMI IJkwezen B.V. Hugo
de Grootplein 1 3314 EG
Dordrecht The
Netherlands

Notified Body Number 122

In accordance with The Council Directive 90/384/EEC on Non-Automatic Weighing Instruments.

Applicant A&D Instruments L TD.
Abingdon Science Park
Abingdon, Oxford OX14 3YS
United Kingdom

In respect of A Class, **II** graduated, self-indicating, electronic, **non-automatic weighing instrument**.
Manufacturer: A&D
Type: FX-EC

Characteristics $n \leq 31000$ divisions
 $200 \text{ g} \leq \text{Max} \leq 6100 \text{ g}$
 $1050 \text{ ct} \leq \text{Max} \leq 30500 \text{ ct}$
 $e = 10d$
 $T \leq -\text{Max}$
Temperature limit: $+ 10 \text{ }^\circ\text{C} \text{ } \cdot / \cdot \text{ } + 30 \text{ }^\circ\text{C}$.

In the description T2170 further essential characteristics are described.

Valid until 27 January 2003

Remarks: This revision replaces the earlier versions of the **EC** type-approval certificates and documentation folders.

Description and documentation The weighing instrument is described in the description number T2170 and documented in the documentation folder number T2170-1, appertaining to this EC type-approval certificate.

Dordrecht, 25 April 1994
NMI IJkwezen B.V.



M. Charité
Director

**Nederlands Meetinstituut Hugo
de Grootplein 1 3314 EG
Dordrecht (NL)** Telephone +31
7833 23 32 Telefax +31
78332309

Nederlands Meetinstituut NV (Registered at the
Chamber of Commerce Delft number 28701)

Subsidiary companies:
NMI Certin BV. (33418)
NMI Van Swinden -boratorium B.V. (28703) NMI
Ukwezen B.V (28700)
NMI Test- en Adviescentrum (TAC) B.V. (28702)

This certificate is issued under the provision that Nederlands
Meetinstituut NV nor its subsidiary companies accept any
liability

Reproduction of the complete certificate is allowed
Parts of the certificate may only be reproduced after written permission