



EC type-approval Certificate

Number **T6716** revision 1
Project number SO15200474
Page 1 of 1

Issued by	NMI Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2009/23/EC, after having established that the measuring instrument meets the applicable requirements of Directive 2009/23/EC, to:
Manufacturer	A&D Instruments Ltd, 24 Blacklands Way Abingdon Business Park Abingdon, Oxfordshire OX14 1DY United Kingdom
Measuring instrument	A Non-automatic weighing instrument Type : HL-3000...WP Further properties are described in the annexes: - Description T6716 revision 1; - Documentation folder T6716-1.
Valid until	16 February 2025
Remark	This revision replaces the earlier version(s), except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
16 February 2015


C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/ndbo/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).

Reproduction of the complete document only is permitted.



1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, shall not be in conflict with the legislation.

1.1 Essential parts

Number	Pages	Description	Remarks
AND-HLWP-BD-H	1	Block diagram (Hardware)	-

The electronics;
 The mechanical assembly with load cell.

EMI protection measures:

- Ferrite on main PCB at the connection point of the power supply and batteries;
- Ferrite on cable between main PCB and load cell.

1.2 Essential characteristics

Accuracy class	III
Maximum capacity	Max ≤ 3 kg
Verification scale interval	$e \geq 1$ g
Weighing range(s)	Single interval
Maximum number of scale intervals (one weighing range)	$n \leq 3000$ divisions
Power supply voltage	230 V AC, 50/60 Hz; 6 x 1,5V batteries.

1.3 Essential shapes

Number	Pages	Description	Remarks
AND-HLWP-EX-1/2	1	Exploded view 1/2	-

The data plate is secured against removal by sealing or will be destroyed when removed.
 Inside the cabinet is a calibration lock, located on the main board.



Description

Number **T6716** revision 1
Project number SO15200474
Page 2 of 3

1.4 Conditional parts

The non automatic weighing instrument is fitted with a levelling device and a level indicator, unless the instrument is installed in a fixed position. The level indicator has a sensitivity of at least 2 mm for a tilt of 2/1000.

1.5 Non-essential parts

- Batteries;
- AC/DC plug-in power supply.

2 Information about the main constituent parts of the non-automatic weighing instrument

2.1 The electronics

2.1.1 Essential parts

Number	Pages	Description	Remarks
AND-HLWP-PZ+-PCB	1	PZ+4333, Lay-out	-
AND-HLWP-PZ+-PL	1	PZ+4333, Parts list	-

2.1.2 Essential characteristics

List of legally relevant functions:

- Determination stability of equilibrium;
- Zero indicator;
- Semi-automatic zero-setting;
- Automatic zero-setting;
- Initial zero-setting;
- Zero-tracking;
- Semi-automatic subtractive tare balancing;
- Indication of stable equilibrium;
- Gravity compensation;
- Calibration / set-up mode via a switch on the main board;
- Acting upon significant faults;
- Checking the display.

2.1.3 Non-essential parts

- Display;
- Keyboard.

2.2 The mechanical assembly with load cell

2.2.1 Essential parts

Number	Pages	Description	Remarks
AND-HLWP-EX-2/2	1	Exploded view 2/2	Load cell assembly.
AND-HLWP-LC-SP	1	HL-WP Series load cell specification	-

2.2.2 Essential characteristics

- $e \geq F_{max}/5000$;
- Excitation power supply 5 V DC.

2.2.3 Essential shapes

Number	Pages	Description	Remarks
AND-HLWP-EX-2/2	1	Exploded view 2/2	-

3 Seals

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the drawing:

Number	Pages	Description	Remarks
AND-HLWP-S	1	Sealing	-

4 Conditions for conformity assessment

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfill the requirements of article 1 of Annex IV of Directive 2009/23/EC.